
Name:	Date:
-------	-------

## Sterile Processing Service



## Across

- **3.** RME is considered sterile unless an event occurs deeming it compromised
- **8.** Any medical equipment to be reused from one patient to another
- 10. Clean to Dirty
- **12.** Must be 12 mil thick
- **15.** Spaulding Classification that is defined as coming into contact with mucous membranes or non-intact skin
- **16.** Occurs regularly to ensure everyone knows what to do in case of a fire
- **17.** Wrapping techique frequently used for smaller trays

## **Down**

- 1. Type of sterilizer used for items that are not heat sensitive
- **2.** Spaulding Classification that is defined as Introduced into the sterile body or blood stream
- 4. Dirty to Clean

- **5.** Pressure of Decontamination
- **6.** Spaulding Classification that is defined as coming into contact with non-intact skin
- **7.** Type of sterilizer that aerates
- **9.** Acronym for Safety Data Sheet
- **11.** Pressure of Preparation Area
- **13.** Wrapping techique frequently used for larger trays **14.** Acronym for Personal
- Protective Equipment